



Certificate No. : L0360-180610

財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

This is to certify that

Chroma Ate Inc.

Physics Calibration Laboratory.

No.68, Huaya 1st Rd., Guishan Dist., Taoyuan City 333, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2005
Accreditation Number : 0360
Originally Accredited : July 01, 1997
Effective Period : July 19, 2018 to July 18, 2021
Accredited Scope : Calibration Field, see described in the Appendix

Chung-Lin Wang

Chung-Lin Wang
President, Taiwan Accreditation Foundation
Date : June 10, 2018



Certificate No. : L0360-180610

財團法人全國認證基金會
Taiwan Accreditation Foundation

Accreditation Number : 0360

Laboratory Head : CHIANG, Ming-Yao

Length

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand / model	document name /no.	mini mum value	units	maxi mum value	units	explanation	value	units
KA2003 digimatic caliper	caliper calibrator Mitutoyo 515-555	caliper calibration procedure (Document No.:E-04-22-037)	0	mm	300	mm	outside diameter (resolution: 0.01 mm)	0.02	mm
			0	mm	300	mm	inside diameter (resolution: 0.01 mm)	0.02	mm

Mass/Force

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand / model	document name /no.	mini mum value	units	maxi mum value	units	explanation	value	units
KC4004 torque calibrator	1.torque arm 100 mm & 250 mm 2.dead weight 100 g to 20 kg	torque calibrator calibration procedure (Document No.:E-04-22-041)	0.49 (5)	N m (kgf cm)	9.8 (100)	N m (kgf cm)	CW	0.61	%
			2.45 (25)	N m (kgf cm)	24.5 (250)	N m (kgf cm)	CW	0.41	%
			0.49 (5)	N m (kgf cm)	9.8 (100)	N m (kgf cm)	CCW	0.61	%
			2.45 (25)	N m (kgf cm)	24.5 (250)	N m (kgf cm)	CCW	0.41	%



Certificate No. : L0360-180610

財團法人全國認證基金會
Taiwan Accreditation Foundation

Note : Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.

Approval Signatory

Approval Signatory	Scope
CHIANG, Ming-Yao	KA2003, KC4004

(Null Below)